

પ્રતિ,

-----  
 -----

વિષય - બાયોગેસ સંશોધન કેન્દ્ર, ગુજરાત વિદ્યાપીઠ, સાદરા, તાજિ- ગાંધીનગર માટે ખરીદવા બાબત, ગુજરાત વિદ્યાપીઠ, નીચે જણાવેલ સોધન માટે મેન્યુફેક્ચરર્સ, ઓથોરાઇઝ ડિલર તથા વહેપારીઓ પાસેથી ભાવ મંગાવે છે. અન્ય વિગત નીચે મુજબ છે.

ક્રમ	સાધનની વિગત	કુલ નંગ	કુલકિંમત (બધા જ કરવેરા તથા ઈન્ટોલેશન સાથે)
01	<p><b><u>ORBITAL SHAKER WITH TEMPERATURE CONTROLLER</u></b></p> <ul style="list-style-type: none"> <li>Brushless Induction Motor with variable frequency drive suitable for continuous operation</li> <li>Step less variable frequency drive ensures gentle shaking start and maintains set speed</li> <li>Powerful fan motor for forced air circulation to maintain uniform conditions inside chamber</li> <li>Counter balanced mechanism for high stability in uneven load of different sized flasks</li> <li>Microprocessor controller with 4" LCD display for display of shaking speed &amp; temperature</li> <li>Time delay for compressor switch on</li> <li>Overload cut off relay for compressor</li> <li>Additional tray to store sample</li> <li>Programs memory</li> <li><b>Temperature range accuracy: 5°C to 60°C, ± 0.5°C</b></li> <li>Internal Volume (Liters): 215 L</li> <li>Platform Size: 580mm x 600mm</li> <li><b>Shaking platform: 250 ml x 33</b></li> <li>Shaking Speed range (RPM): 20 to 250</li> <li>Shaking Amplitude: 25 mm</li> <li>Internal Dimensions W x D x H (mm): 660 x 765 x 650</li> <li>External Dimensions W x D x H (mm): 800 x 1150 x 1300</li> <li>Temperature control: Microprocessor with PT-100 sensor</li> <li>Display: 4" LCD Screen, Large size Display for ease of reading</li> <li>Power Failure Alarm: Audio Visual Alarm</li> <li>Door Open Alarm: Audio Alarm in case door open for over one minute</li> </ul>	01	<p><b><u>Note - Compulsory Aspect of Purchase : Five years Maintenance Contract with Spares (Include in total with Instrument Price )</u></b></p>

	<ul style="list-style-type: none"> <li>• Temperature Variation Alarm: Set Temperature <math>\pm 2^{\circ}\text{C}</math>, Audio Visual Alarm</li> <li>• Illumination: 8 Watts Fluorescent Lamp</li> <li>• Internal Body Material: Stainless Steel - 304 grade</li> <li>• External Body Material: Powder Coated CRCA Steel</li> <li>• Insulation (CFC free polyurethane foam): 70 mm minimum for Body &amp; 80 mm for Door</li> <li>• Noise Level: Less than 65 db</li> <li>• Power: 220-240 volts, 50Hz, Single Phase</li> <li>• <b><u>With Online UPS - 5 KVA (Single phase) with 30 min battery backup :</u></b> DSC based Double conversion online UPS system, <b>Input :</b> 160-280 V AC, Operating frequency range:45-55Hz, 5; <b>Output:</b> 220-230Vac, Operating frequency range:50Hz, Load PF range:0.8 lagging, Output waveform- Pure Sine Wave, General Indication- LCD/LED – Mains failure, Battery high/low, high/low voltage, overload, Forced air cooling, &gt;85% efficiency, <b>With Batteries</b> -18Ah x 15 nos batteries)</li> <li>• <b><u>Compulsory Aspect of Purchase : Five years Maintenance Contract with Spares (Include in total with Instrument Price).Which covers Breakdown and Preventive maintenance visits.</u></b></li> <li>• <b>Training for Users - Onsite to our department laboratory</b></li> </ul>		

ઉપર જણાવેલ ORBITAL SHAKER WITH TEMPERATURE CONTROLLER ના ભાવ આ સાથેના પત્રકમાં વિગત ભરીતા -૨૩/૦૪/૨૦૧૮ ના રોજ સાંજે ૪-૦૦ કલાક સુધીમાં બંધ કરવે ઉપર ORBITAL SHAKER WITH TEMPERATURE CONTROLLER ના ભાવપત્રક (કોટેશન) લખીને કુલસચિવશ્રી ગૂજરાત વિદ્યાપીઠ, અમદાવાદ- ૩૮૦૦૧૪ ના નામે મોકલી આપશો.

(૧) ભાવપત્રક ખોલવાની તારીખ -૨૪-૦૪-૨૦૧૮ ના રોજ મંગળવારે બપોરે ૧-૦૦ કલાકે ખોલવામાં આવશે. જેથી જે તે ભાવ ભરનાર વહેપારી તે સમયે હાજર રહી શકશે.

(૨) જથ્થો વધારો અથવા ઘટાડો કરવાની તથા ઓર્ડર કેન્સલ કરવાની સત્તા ગૂજરાત વિદ્યાપીઠ નેઆધીન રહેશે. તથા કોઈપણ ટેન્ડર સ્વીકારવું કે અસ્વીકાર કરવો કે અંશત સ્વીકારવું તે બાબતનો સંપૂર્ણ અબાધિત અધિકાર ગૂજરાત વિદ્યાપીઠનો રહેશે.

(૩) ઉપરોક્ત ભાવ તમારા લેટર પેડ ઉપર આપવાના રહેશે.

(૪) સાધનની વિગત તથા તેના બ્રાન્ડનેમ સાથે જણાવવાનું રહેશે.

(૫) ભાવપત્રક સાથે સાધનનું બ્રોશર તથા તેને લગતુ સાહિત્ય આપવું. તથા ટેન્ડર ભરનારની માહિતીનું પત્રક ભરીને જોડવાનું રહેશે.

(૬) સાધનની વોરંટી તથા ગેરંટી જણાવવાની રહેશે.

બીડાણ - ટેન્ડર ભરનારની માહિતીનું પત્રક

(ડૉ. રાજેન્દ્ર ખીમાણી)

કુલસચિવ

**GUJARAT VIDYAPITH**  
**NR. INCOMETAX OFFICE, ASHRAM ROAD, AHMEDABAD - 380014**  
**TECHNICAL BID**

**Details of the Tenderer**

<b>Name of the Tender</b>																				

<b>Firm type</b>																				
------------------	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--	--

(Like proprietorship / Partnership / Company / others)

**Address**


**Contact No :**

<b>+Land line No.</b>																				
<b>Fax No.</b>																				
<b>Mobile No.</b>																				
<b>Email address</b>																				

PAN No. \_\_\_\_\_ GST No. \_\_\_\_\_

- Please attach the copy of the GST No with the tender.

Client's List (Please attach separate sheet showing the client list of Universities / Govt. & semi Govt. offices name and addresses of the department where supply the prescribe the tender item with evidence (P.O / certificate of experience)

<b>Sr. No.</b>	<b>Name of Uni/ Govt. &amp; semi Govt office</b>	<b>Year of Supply</b>	<b><u>Name of equipment</u></b>	<b><u>cost of equipment</u></b>

**Signature of the tenderer/agency**  
**Name of the tenderer/agency**

**Seal of the tenderer/agency**

**Place:**

C:\Users\BHAVIKA\Desktop\orbit shaker - inquiry letter.odt